

Effective Date: 16 Sep 2011

PD CARE SURFACE DISINFECTANT/ DECONTAMINANT CLEANER EFFECTIVELY KILLS HUMAN IMMUNODEFICIENCY VIRUS TYPE 1 (HIV), HEPATITIS B VIRUS (HBV) AND HEPATITIS C VIRUS (HCV) ON PRECLEANED ENVIRONMENTAL SURFACES/OBJECTS PREVIOUSLY SOILED WITH BLOOD/BODY FLUIDS IN HEALTHCARE SETTINGS OR OTHER SETTINGS IN WHICH THERE IS AN EXPECTED LIKELIHOOD OF SOILING OF INANIMATE SURFACES/OBJECTS WITH BLOOD/ BODY FLUIDS, AND IN WHICH THE SURFACES/ OBJECTS CAN BE ASSOCIATED WITH THE POTENTIAL FOR TRANSMISSION OF HIV-1, HBV AND HCV.

SPECIAL INSTRUCTIONS FOR CLEANING AND DECONTAMINATION AGAINST HIV-1, HBV AND HCV OF SURFACES/OBJECTS SOILED WITH BLOOD/BODY FLUID:

Personal Protection: Wear appropriate barrier protection such as protective gloves, gowns, masks and eye coverings.

Cleaning Procedure: Blood and other bodily fluids must be thoroughly cleaned from surfaces and objects before disinfection with pdCare Surface Disinfectant/ Decontaminant Cleaner.

Contact Time: A contact time of 2 minutes is required for HIV-1, HBV and HCV. This contact time will not be sufficient for some of the organisms listed on the label. Refer to label for contact times against other organisms.

Infectious Materials Disposal: Cleaning materials used that may contain blood or other bodily fluids must be autoclaved and/or disposed of in accordance with local regulations for infectious materials disposal.

This product is not to be used as a terminal sterilant/ high-level disinfectant on any surface or instrument that (1) is introduced directly into the human body, either into or in contact with the bloodstream or normally sterile areas of the body, or (2) contacts intact mucous membranes but which does not ordinarily penetrate the blood barrier or otherwise enter normally sterile areas of the body. This product may be used to

preclean or decontaminate critical or semi-critical medical devices prior to sterilization/high-level disinfection.

FIRST AID

If in eyes: Hold eye open and rinse slowly and gently with water for 15-20 minutes. Remove contact lenses, if present, after the first 5 minutes, then continue rinsing eye. Call a poison control center or doctor for treatment advice.

If on skin: Take off contaminated clothing. Rinse skin immediately with plenty of water for 15-20 minutes. Call a poison control center or doctor for treatment advice. For Chemical Emergency ONLY (spill, leak, fire, exposure or accident), call CHEMTREC at 1-800-424-9300. For ALL other inquiries about this product, call Customer Service at 1-800-841-1428, Monday through Friday, between 6 a.m. and 4 p.m. Pacific Time.

Have the product container or label with you when calling a poison control center or doctor, or going for treatment.

STORAGE AND DISPOSAL

Do not contaminate water, food, or feed by storage and disposal.


Pesticide Storage: Store in a cool place.

Pesticide Disposal: Wastes resulting from the use of this product may be disposed of according to applicable Federal, State or local procedures.

Container Disposal: Nonrefillable container. Do not reuse or refill this container. Offer for recycling, if available.



+D66207085061009



Ready-To-Use

• TUBERCULOCIDAL** • BACTERICIDAL
• VIRUCIDAL* • FUNGICIDAL

pdCARE™


Surface Disinfectant
Decontaminant Cleaner

ACTIVE INGREDIENTS:
Diisobutylphenoxyethoxyethyl dimethyl benzyl ammonium chloride.....0.28%
Isopropanol17.20%
INERT INGREDIENTS:..... 82.52%
TOTAL100.00%

Precautionary Statements
Hazards to Humans and Domestic Animals.
CAUTION
Harmful if absorbed through skin. Causes moderate eye irritation. Avoid contact with eyes, skin or clothing. Wash hands before eating, drinking, chewing gum, using tobacco or using the toilet. Remove contaminated clothing and wash clothing before reuse.
Physical or Chemical Hazards: Combustible. Do not use or store near heat or open flame.

EPA Reg. No. 46781-6-43100
EPA Est. No. 46781-MI-1

Manufactured for:



Patterson Dental Supply, Inc.
1031 Mendota Heights Rd.,
Saint Paul, MN 55120

KEEP OUT OF REACH OF CHILDREN
CAUTION

See other panel for additional precautionary statements.

Net Contents:
24 fl oz/710 mL
(Net Weight 1.5 lbs./680 g.)

Reorder Number
085-0610

MADE IN U.S.A.

EFFECTIVE AGAINST

- *Mycobacterium tuberculosis* var: bovis (BCG)** • *Staphylococcus aureus* • *Pseudomonas aeruginosa* • *Salmonella enterica*
- *Trichophyton mentagrophytes* • Methicillin Resistant *Staphylococcus aureus* (MRSA) • Vancomycin Resistant *Enterococcus faecalis* (VRE)
- Hepatitis B Virus (HBV)* • Hepatitis C Virus (HCV)*
- Herpes Simplex Virus Types 1 and 2*
- Human Immunodeficiency Virus (HIV-1)*

DESCRIPTION

pdCare™ Surface Disinfectant/Decontaminant Cleaner is a multi-purpose disinfectant/decontaminant cleaner for use on hard, non-porous inanimate surfaces.

It is useful in health care settings such as dental operatories, surgical suites and other critical care areas where control of cross contamination between treated surfaces is important.

pdCare Surface Disinfectant/Decontaminant Cleaner will effectively clean and disinfect, when used as directed, such items as dental unit instrument trays, operatory light switches, exterior surfaces of anesthesia machines and dental curing lights, exterior surfaces of anesthesia machines and respiratory therapy equipment, exterior surfaces of dental operatories, exterior surfaces of endodontic equipment, such as apex locators, pulp testers and motors, stools, floors, walls, handrails, door knobs, bathroom fixtures, exterior toilet surfaces, cabinets, keyboards and telephones that are made of plastic (i.e., polycarbonate, polypropylene, polyvinylchloride, polystyrene), and use sites listed on this label consisting of non-porous vinyl, stainless steel, painted surfaces, Plexiglas®, Formica® and glass.

DIRECTIONS FOR USE

It is a violation of Federal law to use this product in a manner inconsistent with its labeling.

FOR PRECLEANING NON-INSTRUMENT SURFACES PRIOR TO DISINFECTION:

Remove gross filth and heavy soil loads. Spray pdCare Surface Disinfectant/Decontaminant Cleaner directly onto surface. Allow to remain visibly wet for 30 seconds. Wipe surface using a towel.

FOR PRECLEANING INSTRUMENTS PRIOR TO DISINFECTION:

Remove gross filth and heavy soil loads.

For use as a precleaning decontaminant spray: Place instruments in a container. Spray pdCare Surface Disinfectant/Decontaminant Cleaner so as to thoroughly drench all instrument surfaces and allow to remain visibly wet for 30 seconds. Remove and rinse instruments. Follow with appropriate cleaning and disinfection process.

FOR USE AS A VIRUCIDE* ON NON-INSTRUMENT SURFACES, INCLUDING HIV-1, HBV AND HCV APPLICATIONS:

Spray pdCare Surface Disinfectant/Decontaminant Cleaner directly onto precleaned surface, thoroughly wetting area to be treated. Allow surface to remain visibly wet for 2 minutes at room temperature (68°F/20°C).

FOR USE AS A DISINFECTANT ON NON-INSTRUMENT SURFACES:

Spray pdCare Surface Disinfectant/Decontaminant Cleaner directly onto precleaned surface, thoroughly wetting area to be disinfected. Allow surface to remain visibly wet for 3 minutes at room temperature (68°F/20°C) for Tuberculocidal** activity and effectiveness against *Staphylococcus aureus*, *Pseudomonas aeruginosa*, *Trichophyton mentagrophytes* and *Salmonella enterica*.

Wipe surface using a towel or allow to air dry.

FOR USE AS A VIRUCIDE* ON PRECLEANED NON-CRITICAL MEDICAL DEVICES*, INSTRUMENTS AND IMPLEMENTS INCLUDING HIV-1, HBV AND HCV APPLICATIONS:

Instruments must be thoroughly cleaned to remove excess organic debris, rinsed and dried. Thoroughly clean and rinse lumens of hollow instruments. Using either a tray or an ultrasonic unit, immerse instruments in pdCare Surface Disinfectant/Decontaminant Cleaner for 2 minutes at room temperature (68°F/20°C).

FOR USE AS A DISINFECTANT ON PRECLEANED NON-CRITICAL MEDICAL DEVICES*, INSTRUMENTS AND IMPLEMENTS:

Instruments must be thoroughly cleaned to remove excess organic debris, rinsed and dried. Thoroughly clean and rinse lumens of hollow instruments. Using either a tray or an ultrasonic unit, immerse instruments in pdCare Surface Disinfectant/Decontaminant Cleaner for 3 minutes at room temperature (68°F/20°C) for Tuberculocidal** activity and effectiveness against *Staphylococcus aureus*, *Pseudomonas aeruginosa*, *Trichophyton mentagrophytes* and *Salmonella enterica*.

Remove and rinse instruments. Wipe dry prior to use. Discard solution after each use.

* Non-critical medical devices are items that come in contact only with intact skin.

Note: Critical and semi-critical devices must be followed by appropriate terminal sterilization/high-level disinfection process.